

IOtest® CD2-FITC

REF A07743
100 Tests ; 2 mL
20 µL / Test



IOtest
Conjugated Antibody



ENGLISH	Specifications
Specificity	CD2
Clone	39C1.5
Hybridoma	P3-X63-Ag.8.653 x rat spleen cells
Immunogen	PHA stimulated human lymphoblasts
Immunoglobulin	IgG2a
Species	Rat
Source	Ascites
Purification	Ion exchange chromatography
Fluorochrome	Fluorescein isothiocyanate (FITC)
λ Excitation	488 nm
Emission peak	525 nm
Buffer	PBS pH 7.2 plus 2 mg / mL BSA and 0.1% NaN ₃

USE

This fluorochrome-conjugated antibody mixture permits the identification and the counting of cell populations expressing the CD2 antigen present in human biological samples by means of flow cytometry.

PRINCIPLE

This test is based on the ability of specific monoclonal antibodies to bind to the antigenic determinants expressed by leucocytes.

Specific staining of the leucocytes is performed by incubating the sample with the IOtest reagent. The red cells are then removed by lysis and the leucocytes, which are unaffected by this process, are analyzed by flow cytometry.

The flow cytometer measures light diffusion and the fluorescence of cells. It makes possible the delimitation of the population of interest within the electronic window defined on a histogram, which correlates the orthogonal diffusion of light (Side Scatter or SS) and the diffusion of narrow-angle light (Forward Scatter or FS). Other histograms combining two of the different parameters available on the cytometer, can be used as supports in the gating stage depending on the application chosen by the user.

The fluorescence of the delimited cells is analyzed in order to distinguish the positively-stained events from the unstained ones. The results are expressed as a percentage of positive events in relation to all the events acquired by the gating.

EXAMPLES OF CLINICAL APPLICATIONS

The CD2 antigen is present on 80% of peripheral blood lymphocytes in a healthy individual, on 95% of thymocytes and on 100% of T cell lymphocytes (1). A high percentage of "Natural Killer" (NK) cells express them. On the other hand, it is absent in B cells (2, 3). A weak CD2 antigen expression continuum has been described for monocytes, the most positive of which show dendritic cell differentiation characteristics (4).

This reagent permits the characterization and the counting of T lymphocytic cell populations in immune system disorders: immune deficiencies, auto-immune disorders, hypersensitivity reactions, viral infections, restoration of the immune response after bone marrow and/or organ transplantation (5). In malignant blood dyscrasias such as leukaemia and lymphoma, it permits the follow-up and phenotyping of T cell populations (6).

STORAGE AND STABILITY

The conjugated liquid forms must be kept at between 2 and 8°C and protected from light, before and after the vial has been opened.

Stability of closed vial: see expiry date on vial.

Stability of opened vial: the reagent is stable for 90 days.

PRECAUTIONS

1. Do not use the reagent beyond the expiry date.
2. Do not freeze.
3. Let it come to room temperature (18 – 25°C) before use.
4. Minimize exposure to light.
5. Avoid microbial contamination of the reagents, or false results may occur.
6. Antibody solutions containing sodium azide (NaN₃) should be handled with care. Do not take internally and avoid all contact with the skin, mucosa and eyes.
Moreover, in an acid medium, sodium azide can form the potentially dangerous hydrazoic acid. If it needs to be disposed of, it is recommended that the reagent be diluted in a large volume of water before pouring it into the drainage system so as to avoid the accumulation of sodium azide in metal pipes and to prevent the risk of explosion.
7. All blood samples must be considered as potentially infectious and must be handled with care (in particular: the wearing of protective gloves, gowns and goggles).
8. Never pipette by mouth and avoid all contact of the samples with the skin, mucosa and eyes.
9. Blood tubes and disposable material used for handling should be disposed of in ad hoc containers intended for incineration.

SAMPLES

Venous blood or bone marrow samples must be taken using sterile tubes containing an EDTA salt as the anticoagulant. The use of other anticoagulants is not recommended.

The samples should be kept at room temperature (18 – 25°C) and not shaken. The samples should be homogenized by gentle agitation prior to taking the test sample.

The samples must be analyzed within 24 hours of taking them.

METHODOLOGY

NECESSARY MATERIAL NOT SUPPLIED

- Sampling tubes and material necessary for sampling.
- Automatic pipettes with disposable tips for 20, 100 and 500 µL.
- Plastic haemolysis tubes.
- Calibration beads: Flow-Set™ Fluorospheres (Ref. 6607007).
- Red cell lysis reagent with washing stage after lysis. For example: VersaLyse (Ref. A09777).
- Leucocyte fixation reagent. For example : IOtest 3 Fixative Solution (Ref. A07800).
- Isotypic control: FITC-conjugated rat IgG2a.
- Buffer (PBS: 0.01 M sodium phosphate; 0.145 M sodium chloride; pH 7.2).
- Centrifuge.
- Automatic agitator (Vortex type).

- Flow cytometer.

PROCEDURE

NOTE: The procedure below is valid for standard applications. Sample and/or VersaLyse volumes for certain Beckman Coulter applications may be different. If such is the case, follow the instructions on the application's technical leaflet. For each sample analyzed, in addition to the test tube, one control tube is required in which the cells are mixed in the presence of an Isotypic control.

1. Add 20 µL of specific IOtest conjugated antibody to each test tube, and to each control tube, the necessary amount of the isotypic control.
2. Add 100 µL of the test sample to both tubes. Vortex the tubes gently.
3. Incubate for 15 to 20 minutes at room temperature (18 – 25°C), protected from light.
4. Then perform lysis of the red cells, if necessary, by following the recommendations of the lysis reagent used. As an example, if you wish to use VersaLyse (Ref. A09777), refer to the leaflet and follow preferably the procedure called "with concomitant fixation", which consists of adding 1 mL of the "Fix-and-Lyse" mixture prepared extemporaneously. Vortex immediately for one second and incubate for 10 minutes at room temperature, protected from light.

If the sample does not contain red cells, add 2 mL of PBS.

5. Centrifuge for 5 minutes at 150 x g at room temperature.
6. Remove the supernatant by aspiration.
7. Resuspend the cell pellet using 3 mL of PBS.
8. Repeat step 5.
9. Remove the supernatant by aspiration and resuspend the cell pellet using:
 - 0.5 mL or 1 mL of PBS plus 0.1% of formaldehyde if the preparations are to be kept for more than 2 hours and less than 24 hours. (A 0.1% formaldehyde PBS can be obtained by diluting 12.5 µL of the IOtest 3 Fixative Solution (Ref. A07800) at its 10X concentration in 1 mL of PBS).
 - 0.5 mL or 1 mL of PBS without formaldehyde, if the preparations are to be analyzed within 2 hours.

NOTE: In all cases, keep the preparations between 2 and 8°C and protected from light.

PERFORMANCE

SPECIFICITY

Monoclonal antibody 39C1.5 inhibits rosette formation with sheep red cells (7). It was assigned to CD2 during the 2nd HLDA Workshop on Human Leucocyte Differentiation Antigens, held in Boston, United States, in 1984 (WS Code: 144, Section T) (8).

LINEARITY

To test the linearity of staining of this reagent, a positive cell line (JURKAT) and a negative cell line (RAMOS) were mixed in different proportions with a constant final number of cells, so that the positive line/negative line ratio of the mixture ranged from 0 to 100%.

Aliquots were stained using the procedure described above and linear regression between the expected values and the observed values was calculated. The parameters of the equation of the linear regression may be used to determine the linearity as well as the range of measurement for each specificity.

Specificity	Linear regression	Linearity (R ²)	Range (%)
CD2	$Y = 0.97 X + 1.17$	0.999	2 – 97

EXPECTED VALUES

Each laboratory must compile a list of reference values based upon a group of healthy donors from the local population. This must be done by taking age, sex and ethnic group into account, as well as any other potential regional differences.

In our laboratories, the whole blood of 50 healthy adults was treated using the reagent described above. The results obtained for the counting of the positive events of interest with this reagent are given in the tables below:

Lymphocytes	Number	Mean (%)	SD	CV (%)
CD2 ⁺	50	82.0	6.6	8

INTRA-LABORATORY REPRODUCIBILITY

On the same day and using the same cytometer, 12 measurements of the percentage of positive cells were carried out on a target population (peripheral blood lymphocytes). The results obtained are summarized in the following table:

Positive-Target	Number	Mean (%)	SD	VK (%)
CD2 ⁺ Lymphocytes	12	83.7	0.62	0.7

INTER-LABORATORY REPRODUCIBILITY

On the same day and for the same population (peripheral blood lymphocytes), 12 measurements of the percentage of positive cells were carried out by two technicians and the preparations analyzed using two different cytometers. The results obtained are summarized in the following tables:

Cytometer n° 1:

Positive-Target	Number	Mean (%)	SD	VK (%)
CD2 ⁺ Lymphocytes	12	83.7	0.62	0.7

Cytometer n° 2:

Positive-Target	Number	Mean (%)	SD	VK (%)
CD2 ⁺ Lymphocytes	12	86.7	0.51	0.6

LIMITATIONS OF THE TECHNIQUE

1. Flow cytometry may produce false results if the cytometer has not been aligned perfectly, if fluorescence leaks have not been correctly compensated for and if the regions have not been carefully positioned.
2. It is preferable to use a lysis technique with washing as this reagent has not been optimized for "without washing" lysis techniques.
3. Accurate and reproducible results will be obtained as long as the procedures used

are in accordance with the technical insert leaflet and compatible with good laboratory practices.

4. The conjugated antibody of this reagent is calibrated so as to offer the best specific signal/non-specific signal ratio. Therefore, it is important to adhere to the reagent volume/number of cells ratio in every test.
5. In the case of a hyperleucocytosis, dilute the blood in PBS so as to obtain a value of approximately 5×10^9 leucocytes/L.
6. In certain disease states, such as severe renal failure or haemoglobinopathies, lysis of red cells may be slow, incomplete or even impossible. In this case, it is recommended to isolate mononucleated cells using a density gradient (Ficoll, for example) prior to staining.

MISCELLANEOUS

See the Appendix for examples and references.

TRADEMARKS

The Beckman Coulter logo, COULTER, EPICS, EXPO, Flow-Set, IOTest, System II, XL are registered trademarks of Beckman Coulter Inc.

BD FACScan is a registered trademark of BD Biosciences and Company.

MANUFACTURED BY:

IMMUNOTECH S.A.
a Beckman Coulter Company
130 avenue de Lattre de Tassigny
B.P. 177 – 13276 Marseille Cedex 9
France
Customer Services: (33) 4 91 17 27 27

www.beckmancoulter.com




APPENDIX TO REF A07743


EXAMPLES

The graphs below are monoparametric representations (Count versus Fluorescence Intensity) of lysed normal whole blood sample. Staining is with IOTest CD2-FITC Conjugated Antibody (Ref. A07743). Gate is on lymphocytes.

Acquisition and analysis are performed with a COULTER® EPICS® XL™ flow cytometer equipped with System II™ software.



Acquisition is with a Becton Dickinson FACScan™ flow cytometer.



REFERENCES

1. Bierber, B.E., Bogart, R.E., Wolff, I.H., Burkakoff, S.J., "Functional analysis of CD2 Mab reactivity", 1989, *Leucocyte Typing IV, White Cell Differentiation Antigens*. W. Knapp, et al., Eds., Oxford University Press, p. 274-277.
2. Sewel, W.A., Brown, M.H., Dunne, J., Owen, M.J., Crumpton, M.J., "Molecular cloning of the human T lymphocyte surface CD2 (T11) antigen", 1986, *Proc.Natl. Acad.Sci. USA*, 83, 8718-8722.
3. Hannel, I., Erkeller-Yuksel, F., Lydyard, P., Deneys, V., De Bruyère, M., "Developmental and maturational changes in human blood lymphocyte subpopulations", 1992, *Immunol. Today*, 13, 215.
4. Crawford, K., Gabuzda D., Pantazopoulos, V., Xu, J., Clement, C., Reinherz, E., Alper, C., "Circulating CD2⁺ Monocytes are Dendritic Cells", 1999, *J. Immunology*, 163, 5920-5928.
5. Sprent, J., "T lymphocytes and the thymus", 1989, in *Fundamental Immunology*, 2nd edition, Paul, W.E., Ed., Raven Press, 69-93.
6. Vaikus, L., Ball, E.D., Foon, K. A., "Immune markers in hematologic disease", 1991, *Crit. Rev. Onco/Hematol.*, 11, 267.
7. Olive, D., Ragueneau, M., Cerdan, C., Dubreuil, P., Lopez, M., Mawas, C., "Anti-CD2 (Sheep red blood cell receptor) monoclonal antibodies and T-cell activation. Pairs of anti-T11-1 and T11-2 (CD2 subgroups) are strongly mitogenic for T cells in presence of 12-O tetradecnoylphorbol 13 acetate", 1986, *Eur. J. Immunol.*, 16, 1063-1068.
8. Haynes, B.F., "Summary of T cell studies", 1986, in *Leukocyte Typing II, White Cell Differentiation Antigen*, Rheinherz, E.L., et al., Eds., Springer-Verlag, p. 3-30.